

GC907-23 Performed Tape Sealant Sample Holder
 Tests in accordance with ASTM C907

TestResources GC907-23 Performed Tape Sealant Sample Holder measures tensile adhesive strength in performed tape sealant in strict accordance with ASTM Standard C907. The preformed tape sealant is compressed between two aluminum disks of the fixture during testing. ASTM C907 can also provide an indication of the cohesive nature and adhesive bonding properties of the preformed tape sealant after testing.

GC907-23 includes one upper and one lower fixture. The fixture has 15.9mm female cup adapters with a hole for an 8mm diameter cross pin (additional machine adapters available). Purchase of GC907-23 includes one pair of extra sample disks.

